

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Masterclave 10 Media Preparator

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and fax this form to us on **01274 531197**. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

**The following information represents the ideal requirement.
Please contact us IMMEDIATELY if your intended location does not match this specification.**



Access Requirements		<input type="checkbox"/>																				
The unit can fit through a door opening of 700mm (27.5") but check size of lifts, stairs etc.																						
Space Requirements																						
The weight of the equipment is 85kg.																						
The unit should be sited as close as possible (preferably to the right) of the pourer stacker																						
External Dimensions																						
<table border="1" style="width: 100%; border-collapse: collapse; text-align: center;"> <thead> <tr> <th style="padding: 2px;">Depth</th> <th style="padding: 2px;">Length</th> <th style="padding: 2px;">Height</th> </tr> </thead> <tbody> <tr> <td style="padding: 2px;"><i>mm</i></td> <td style="padding: 2px;"><i>mm</i></td> <td style="padding: 2px;"><i>mm</i></td> </tr> <tr> <td style="padding: 2px;">550</td> <td style="padding: 2px;">440</td> <td style="padding: 2px;">650</td> </tr> </tbody> </table>	Depth	Length	Height	<i>mm</i>	<i>mm</i>	<i>mm</i>	550	440	650	<input type="checkbox"/>												
Depth	Length	Height																				
<i>mm</i>	<i>mm</i>	<i>mm</i>																				
550	440	650																				
Mains Requirements																						
Electricity Supply	Socket	<input type="checkbox"/>																				
230 +/-10% V, single phase 50hz	Three pin 16 amp																					
Water System Requirements		<input type="checkbox"/>																				
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="padding: 2px;">Input (Tap water)</th> <th style="padding: 2px;"></th> <th style="padding: 2px;">Output</th> </tr> </thead> <tbody> <tr> <td style="padding: 2px;">Minimum flow</td> <td style="padding: 2px;">5 l/min</td> <td rowspan="2" style="padding: 2px;">Copper or PVC pipe that may hold high temperatures (130°C)</td> </tr> <tr> <td style="padding: 2px;">Minimum pressure</td> <td style="padding: 2px;">1 bar</td> </tr> <tr> <td style="padding: 2px;">Maximum pressure</td> <td style="padding: 2px;">4 bar</td> <td style="padding: 2px;">Pipe minimum diameter: 40mm</td> </tr> <tr> <td style="padding: 2px;">Minimum temperature</td> <td style="padding: 2px;">5°C</td> <td style="padding: 2px;">No pressure at output</td> </tr> <tr> <td style="padding: 2px;">Maximum temperature</td> <td style="padding: 2px;">20°C</td> <td></td> </tr> <tr> <td style="padding: 2px;">Maximum water hardness</td> <td style="padding: 2px;">16°TH</td> <td></td> </tr> </tbody> </table>	Input (Tap water)		Output	Minimum flow	5 l/min	Copper or PVC pipe that may hold high temperatures (130°C)	Minimum pressure	1 bar	Maximum pressure	4 bar	Pipe minimum diameter: 40mm	Minimum temperature	5°C	No pressure at output	Maximum temperature	20°C		Maximum water hardness	16°TH			
Input (Tap water)		Output																				
Minimum flow	5 l/min	Copper or PVC pipe that may hold high temperatures (130°C)																				
Minimum pressure	1 bar																					
Maximum pressure	4 bar	Pipe minimum diameter: 40mm																				
Minimum temperature	5°C	No pressure at output																				
Maximum temperature	20°C																					
Maximum water hardness	16°TH																					
Connection		<input type="checkbox"/>																				
¾ " bsp male (washer connection) fitting or a ½ " (15mm) barb. In both cases, an isolation tap is recommended.																						
Drain		<input type="checkbox"/>																				
1 ¾ " (30mm) or 1 ½ " (38mm) waste trap or a sufficiently large bore waste pipe to accommodate one 1" (25mm) hose.																						

Validation is highly recommended for each load to be utilised in order to:

- ensure that the required temperatures and times throughout the load are achieved
and
- determine cycle settings.

As Don Whitley Scientific is **UKAS accredited**, we can offer you validation and calibration of your heat sterilisation equipment.

If you would like to discuss this, please call us on 01274 595728.

Delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, **a charge may be made.**

It is essential that this form is completed and returned to avoid delay to your installation.

THANK YOU FOR THINKING WHITLEY

Signature

Title

Print Name

Establishment