

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Whitley H55 Hypoxystation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete, and email this form to us at service@dwscientific.co.uk. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal installation requirements.

Please contact us IMMEDIATELY if your intended location does not match this specification.

				\checkmark
Access Requirements				
For access, the dimensions be	low should be considered when checking t	he size of doorways, I	ifts, stairs, etc	
Space Requirements				
but please ensure that sufficier	s 200 kg (or 210kg if fitted with the Whitley nt floor space is allocated. If you chose to be will have to be supplied for the left-hand p	ench mount the works	station, please contact a member of the	
Width (mm)	Depth (mm)	Height (mm)]	
2510	720 (or 810 if HEPA option fitted)	710		
* Please Note: In addition to the dimensions	stated above, allow a localised protrusion	of 90mm at rear to acc	commodate condenser bottle, bracket,	

In addition to the dimensions stated above, allow a localised protrusion of 90mm at rear to accommodate condenser bottle, bracket, gas and electrical supplies. Wherever the unit is situated, a further minimum clearance of 500mm is required above the unit and a minimum clearance of 200mm at the right-hand side of the unit is required for user/service access.

The H55 will be shipped in two parts (left hand workstation and right-hand workstation) to facilitate delivery to your laboratory (negotiating lifts, corridors, doorways, etc). Each half of the workstation measures 1255mm wide x 720mm (or 810mm if HEPA option fitted) deep x 710mm high.

Gas Requirements

The incoming gas supplies must be terminated near the right-hand side of the main chamber and fitted with leak-proof taps and pressure gauges.

The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.

Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.

Gas Type	Connection Details	Cylinder Regulator Required	Regulator Outlet Range	Flow Rate
CO ₂	1/4 BSP male fitting or connection for 6mm Polyurethane Tubing CO₂ Regulator – Two Stage – order Code A01747	Two Stage	4-6 bar (60-90 psi)	Minimum 10 litres per minute (Dynamic)
Air	1/4 BSP male fitting or connection for 6mm Polyurethane Tubing Air Regulator – Two Stage – order Code – A01554	Two Stage	4-6 bar (60-90 psi)	Minimum 10 litres per minute (Dynamic)
				Continued overleaf/

DWS/SSP0005 Revision Level: 05

Nitrogen	1/4 BSP male fitting or connection for 8mm Polyurethan Tubing Nitrogen Regulator – Two Stage – order Code A01748	Two Stage	4-6 bar (60-90 psi)	Minimum 150 litres per minute (Dynamic)	
Suitable Connec	ction Types (to affix to gas ou	lets on bottle/wall):			
	Push in con (Fittings of ch	nection pice for DWS. Supplied wi	th DWS spares kit).		
OR					
	Push on col (Customers' o	nnection wn preference. Not suppli	ed by DWS).		
Mains Requirem	ents				
Electricity Supp	ly	Wall Socket			
240 Volts	240 Volts 1 x Three Pin, 13 Amp. Minimum 6A rating				
Other Considera		da			
Do not site the instrument near draughts caused by windows, doors or air conditioning systems Remember, if you do not have the required regulators, you can order them from Don Whitley Scientific: Carbon Dioxide Regulator – order Code A01747 Air Regulator – Two Stage – order Code A01554 Nitrogen Regulator – Two Stage – order Code A01748					
Decontamination and Removal					
If an existing unit is being taken in part exchange or is being removed from the laboratory, it must be de-contaminated before DWS staff handle the unit. A certificate or signed letter confirming the unit has been decontaminated must be given to our engineer.					
There is a £400 fee for DWS to remove an existing unit from site. Please tick to accept this charge and an invoice will be provided.					
In the UK, delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, a charge may be made . Export customers, please refer to your local distributor.					

It is essential that this form is completed and returned, to avoid delay to your installation.

THANK YOU FOR THINKING WHITLEY

Signature	Title	
Print Name	Establishment	

DWS/SSP0005 Revision Level: 05